### § 866.3065

# §866.3065 Bordetella spp. serological reagents.

(a) Identification. Bordetella spp. serological reagents are devices that consist of antigens and antisera, including antisera conjugated with a fluorescent dye, used in serological tests to identify Bordetella spp. from cultured isolates or directly from clinical specimens. The identification aids in the diagnosis of diseases caused by bacteria belonging to the genus Bordetella and provides epidemiological information on these diseases. Bordetella spp. cause whooping cough (Bordetella pertussis) and other similarly contagious and acute respiratory infections characterized by pneumonitis (inflammation of the lungs).

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 54 FR 25046, June 12, 1989; 66 FR 38791, July 25, 2001]

# §866.3085 Brucella spp. serological reagents.

(a) Identification. Brucella spp. serological reagents are devices that consist of antigens and antisera used for serological identification of Brucella spp. from cultured isolates derived from clinical specimens or to identify antibodies to Brucella spp. in serum. Additionally, some of these reagents consist of antisera conjugated with a fluorescent dye (immunofluorescent reagents) used to identify Brucella spp. directly from clinical specimens or cultured isolates derived from clinical specimens. The identification aids in the diagnosis of brucellosis (e.g., undulant fever, Malta fever) caused by bacteria belonging to the genus Brucella and provides epidemiological information on diseases caused by these microorganisms.

(b) Classification. Class II (special controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 63 FR 59226, Nov. 3, 1998]

#### §866.3110 Campylobacter fetus serological reagents.

(a) Identification. Campylobacter fetus serological reagents are devices that consist of antisera conjugated with a fluorescent dye used to identify Campylobacter fetus from clinical specimens or cultured isolates derived from clinical specimens. The identification aids in the diagnosis of diseases caused by this bacterium and provides epidemiological information on these diseases. Campylobacter fetus is a frequent cause of abortion in sheep and cattle and is sometimes responsible for endocarditis (inflammation of certain membranes of the heart) and enteritis (inflammation of the intestines) in humans.

(b) Classification. Class I (general controls).

# $\$\,866.3120$ Chlamydia serological reagents.

(a) Identification. Chlamydia serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to chlamydia in serum. Additionally, some of these reagents consist of chlamydia antisera conjugated with a fluorescent dye used to identify chlamydia directly from clinical specimens or cultured isolates derived from clinical specimens. The identification aids in the diagnosis of disease caused by bacteria belonging to the genus Chlamydia and provides epidemiological information on these diseases. Chlamydia are the causative agents of psittacosis (a form of pneumonia). lymphogranuloma venereum (a venereal disease), and trachoma (a chronic disease of the eve and evelid).

(b) Classification. Class I (general controls).

## §866.3125 Citrobacter spp. serological reagents.

(a) Identification. Citrobacter spp. serological reagents are devices that consist of antigens and antisera used in serological tests to identify Citrobacter spp. from cultured isolates derived from clinical specimens. The identification aids in the diagnosis of disease caused by bacteria belonging to the genus Citrobacter and provides epidemiological information on diseases

### Food and Drug Administration, HHS

caused by these microorganisms. *Citrobacter* spp. have occasionally been associated with urinary tract infections.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to the limitations in §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 54 FR 25046, June 12, 1989; 66 FR 38791, July 25, 2001]

### §866.3135 Coccidioides immitis serological reagents.

(a) Identification. Coccidioides immitis serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to Coccidioides immitis in serum. The identification aids in the diagnosis of coccidioidomycosis caused by a fungus belonging to the genus Coccidioides and provides epidemiological information on diseases caused by this microorganism. An infection with Coccidioides immitis produces symptoms varying in severity from those accompanying the common cold to those of influenza.

(b) Classification. Class II (special controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 63 FR 59226, Nov. 3, 1998]

### §866.3140 Corynebacterium spp. serological reagents.

(a) Identification. Corynebacterium spp. serological reagents are devices that consist of antisera conjugated with a fluorescent dye used to identify Corynebacterium spp. from clinical specimens. The identification aids in the diagnosis of disease caused by bacbelonging to the Corynebacterium and provides epidemiological information on diseases caused by these microorganisms. The principal human pathogen of this Corynebacterium diphtheriae, diphtheria. However, many other types of corynebacteria form part of the normal flora of the human respiratory tract, other mucus membranes, and skin, and are either nonpathogenic or have an uncertain role.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

 $[47\ FR\ 50823,\ Nov.\ 9,\ 1982,\ as\ amended\ at\ 65\ FR\ 2311,\ Jan.\ 14,\ 2000]$ 

# §866.3145 Coxsackievirus serological reagents.

(a) Identification. Coxsackievirus serological reagents are devices that consist of antigens and antisera used in serological tests to identify antibodies to coxsackievirus in serum. Additionally, some of these reagents consist of coxsackievirus antisera conjugated with a fluorescent dve that are used to identify coxsackievirus from clinical specimens or from tissue culture isolates derived from clinical specimens. The identification aids in the diagnosis of coxsackievirus infections and provides epidemiological information on diseases caused by these viruses. Coxsackieviruses produce a variety of infections, including common colds, meningitis (inflammation of brain and spinal cord membranes), herpangina (brief fever accompanied by ulcerated of lesions the throat), and myopericarditis (inflammation of heart tissue).

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter subject to §866.9.

[47 FR 50823, Nov. 9, 1982, as amended at 65 FR 2311, Jan. 14, 2000]

## §866.3165 Cryptococcus neoformans serological reagents.

Identification. Cryptococcus neoformans serological reagents are devices that consist of antigens used in serological tests to identify antibodies to Cryptococcus neoformans in serum. Additionally, some of these reagents consist of antisera conjugated with a fluorescent dye (immunofluorescent reagents) and are used to identify Cryptococcus neoformans directly from clinical specimens or from cultured isolates derived from clinical specimens. The identification aids in the diagnosis of cryptococcosis and provides epidemiological information on this